

PRODUCT BULLETIN

Product(s):	Fabius Tiro
Bulletin Number:	162
Release Date:	January 12, 2004
Pages:	2

Subject: Fabius Tiro COSY breathing system change


Effective immediately, all new production Fabius Tiro systems are being shipped with a COSY that is compatible with the CLIC absorber system.

General Overview


The COSY change is a resolution of the compatibility issue with the CLIC absorber system. As a result of this change, the part number for the COSY has changed.

	<u>US</u>	<u>Canada</u>
New COSY Part #	4118378	4118379

With the new COSY, the bag location has moved slightly.



Old Cosy



New Cosy

As a result of the bag location change, the new COSY will require a different bag arm mount.

Swivel Bag Arm Part # (new COSY only)	4118380	\$ 450
--	----------------	---------------

Please note that the old COSY is only compatible with the old bag arm mount, and the new COSY is only compatible with the new bag arm mount.

Barry Feldman will be updating the quote template and posting it to the Anesthesia Marketing Library by January 16, 2004 to reflect this change.

Field Install Plan

There will be no upgrade to the current Fabius Tiro customers unless they are planning on purchasing the Draeger CLIC. For those customers, an upgrade program to the new COSY will be established. The details of this program are not yet finalized, but are expected by the beginning of the first quarter.

Feel free to contact Christine MacMillan at ext. 2303 or macmillanc@draegermed.com if you have additional questions.